



Raz-AT

Attendant Tilt

Rehab Shower Commode

The RAZ-AT (Attendant Tilt) rehab shower commode chair tilts up to 40° to provide pressure reduction and maximum comfort. The unique frame design utilizes lockable gas springs that help lift the seat to an upright position, making the RAZ-AT one of the easiest tilting chairs to operate. The easy-to-clean, stainless steel frame fits over most toilets and is designed with unobstructed side and front access for the user or attendant.



MMS
MEDICAL

www.mmsmedical.ie

FEATURES

Extensive range of options available



ALSO AVAILABLE

- 16" and 20" (40 cm and 51 cm) widths available
- Recline module
- Shown with optional laterals and H-Strap

SPECIFICATIONS	Raz-AT (Attendant Tilt)
WEIGHT CAPACITY	350 lbs (160 kg)
FRAME WIDTH (seat and back frame)	Standard: 18" (46 cm)
OVERALL WIDTH (with standard armpad)	22.5" (57 cm) with Standard Arm Pad
OVERALL LENGTH (with length (with footrests)	39" (99 cm)
SEAT WIDTH (commode seat)	18" (46 cm)
SEAT DEPTH (front of back post to front of seat)	16" or 19" (40 or 48 cm)
CLEARANCE HEIGHT UNDER SEAT	17" - 21" (43 - 53cm) with pan brackets 17.75" - 21.75" in 1" increments (45 - 55 cm)
ARMREST HEIGHT ABOVE SEAT (20" Adjusta-Back)	7.5" - 10.5" in ½" increments (19 - 27 cm)
SEAT-TO-FLOOR HEIGHT	21.25" - 25.25" in 1" increments (54 - 64 cm)
PRODUCT WEIGHT	61 lbs (28 kg)
FRAME AND HARDWARE MATERIAL	Stainless Steel (Lifetime warranty against corrosion)

All measurements taken with standard molded seat and standard arm pad.

Two-year limited warranty covers the frame against manufacturing defects. A one-year limited warranty covers all other components, subject to normal wear and tear exclusions.

Specifications, and options are subject to change without prior notice.

MANUFACTURED BY: RAZ Design Inc. 22 Howden Road, Toronto, Ontario Canada, M1R 3C7

MMS Medical, 51 Eastgate Drive, Little Island, Cork. Tel : 021 4618000 Email: info@mmsmedical.ie

MMS Medical, Unit 81 Western Parkway, Business Park, Lower Ballymount Road, Dublin 12. Tel: 01 4500582

Rev. Date: 2017-02-27