

APPLICATION
*HIGH RISK USER
WITH POSTURE CONTROL

Sizes: Up to 20"x18".

Larger sizes available POA.

Mercury 300 Xtreme

QM300X



MMS Medical 51 East Gate Drive, Little Island, Cork. T: 021 461 8000

T: 021 461 8000 F: 021 461 8099 E: sales@mmsmedical.ie MMS Medical
Unit 81,
Western Parkway Business Park,
Lower Ballymount Road,
Dublin 12.
T: 01 450 0582
F: 01 456 5889
E: sales@mmsmedical.ie







www.mmsmedical.ie

Mercury 300 Xtreme

Description: A highly contoured seating component designed for high risk clients with particular

attention given to the I.T. area and positional requirements. Combining a highly durable base for control and security, integrated with memory foam and flow-gel

technology resulting in an extremely effective high risk component.

Sizes: Up to 20"x18". Larger sizes are available upon request. Height is approximately 3½"

in the seating area, bariatric sizes may be thicker to accommodate higher loadings.

Cover: The cover is constructed from a vapour permeable clinically proven material,

chosen for its medical performance, durability and ease of use. It is fitted with a heavy duty concealed zip with an easy to reach and use bright yellow slider in a concealed envelope. A non slip base and inner vapour permeable liner are added for security and hygiene. All relevant safety, clinical and maintenance information

is boldly printed on the rear.

Risk Category*: High, with posture control

Foam Insert: A low profile seating component using a cold cure moulded and laminated

construction, incorporating memory foam and a flow gel insert, chosen for their individual properties of posture control, durability, clinical performance and long

term comfort.

Weight Guide[†]: For maximum comfort, durability and performance the weight

guide on a 17"x17" cushion is 17-18 stone (108-114kg)

Support Pressure: 1½ lbs per square inch at 50% compression or deflection

Linear Load Limit: 31 stone (196kg) on a typical 17"x17" cushion

* Risk categories are given as a guide and rely on your judgement and experience.

[†] Weight Guides are a general indication of an ideal user weight to give optimum durability, comfort and performance; it is not a definitive or maximum value. (This product has been tested to 325lb (146kg) without failure).

Instructions for Use and Cleaning

DESIGN

The device has been designed to provide the correct levels of pressure reduction, support and comfort which are needed and according to the design the correct amount of postural assistance.

PLACEMENT

- 1. Before using the cushion for the first time, please remove the outer packaging if present and check that the chair seat has no objects or projections which may damage the new cushion.
- 2. Place the device onto the chair ensuring that our label is on the base of the device (ie. do not sit or lean on the label side).

CLEANING

The cushion has removable covers and can be cleaned according to the material type. If possible sponge off any spillage or contamination as soon as possible with hot water and soap. Do not remove the special transparent liner (if fitted) but sponge it down.

DO NOT PLACE THE WHOLE DEVICE DIRECTLY INTO A WASHING MACHINE

DOMESTIC CLEANING

All products have an approved cleaning label which you should follow to ensure durability and cleanliness of the product.

ALL WATERPROOF COVERS wipe over with hot water and soap, a mild hypochlorite solution can also be used (Milton) if desired. Vapour permeable material, towelling and spacer covers can be machine washed at 60 degrees and tumbled dry on warm only. Leave to air dry before refitting. Do not use furniture polish on the cover.

CLEANING INSTITUTIONAL VAPOUR PERMEABLE MATERIAL

Can be washed at 60 degrees Celsius or dry cleaned.

Wipe over with full strength Sodium Hypochlorite or Sodium Dichloroisocyanurate and maintain surface contact for 2-5 minutes. Wipe off and rinse with clean water and dry.

DO NOT USE PHENOL BASED PRODUCTS ON ANY WATERPROOF COVER

If MRSA is suspected or confirmed please contact your Infection Control Department for advice.